UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/550,137	09/25/2006	Mats Inganas	HO-P02773US1	1324
	7590 01/22/201 & JAWORSKI, LLP	EXAMINER		
1301 MCKINN			YANG, NELSON C	
SUITE 5100 HOUSTON, TX 77010-3095			ART UNIT	PAPER NUMBER
			1641	
			NOTIFICATION DATE	DELIVERY MODE
			01/22/2010	ELECTRONIC

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Notice of the Office communication was sent electronically on above-indicated "Notification Date" to the following e-mail address(es):

nstacey@fulbright.com twrye@fulbright.com hoipdocket@fulbright.com

	Application No.	Applicant(s)
	10/550,137	INGANAS ET AL.
Office Action Summary	Examiner	Art Unit
	Nelson Yang	1641
The MAILING DATE of this communication ap Period for Reply	opears on the cover sheet with the o	correspondence address
A SHORTENED STATUTORY PERIOD FOR REPL WHICHEVER IS LONGER, FROM THE MAILING I - Extensions of time may be available under the provisions of 37 CFR 1 after SIX (6) MONTHS from the mailing date of this communication. - If NO period for reply is specified above, the maximum statutory period - Failure to reply within the set or extended period for reply will, by statul Any reply received by the Office later than three months after the maili earned patent term adjustment. See 37 CFR 1.704(b).	DATE OF THIS COMMUNICATION .136(a). In no event, however, may a reply be tind d will apply and will expire SIX (6) MONTHS from te, cause the application to become ABANDONE	N. mely filed the mailing date of this communication. ED (35 U.S.C. § 133).
Status		
Responsive to communication(s) filed on <u>03 I</u> This action is FINAL . 2b) ☑ This action is FINAL . Since this application is in condition for allowed closed in accordance with the practice under	is action is non-final. ance except for formal matters, pro	
Disposition of Claims		
4) Claim(s) 1-17 and 19-29 is/are pending in the 4a) Of the above claim(s) 3-6,22 and 23 is/are 5) Claim(s) is/are allowed. 6) Claim(s) 1,2,7-17,19-21 and 24-29 is/are rejee 7) Claim(s) is/are objected to. 8) Claim(s) are subject to restriction and/ Application Papers 9) The specification is objected to by the Examin	e withdrawn from consideration. ected. for election requirement.	
10) ☐ The drawing(s) filed on 21 September 2005 is Applicant may not request that any objection to the Replacement drawing sheet(s) including the correct 11) ☐ The oath or declaration is objected to by the E	s/are: a)⊠ accepted or b)⊡ object e drawing(s) be held in abeyance. Se ction is required if the drawing(s) is ob	e 37 CFR 1.85(a). ejected to. See 37 CFR 1.121(d).
Priority under 35 U.S.C. § 119		
 12) Acknowledgment is made of a claim for foreig a) All b) Some * c) None of: 1. Certified copies of the priority documer 2. Certified copies of the priority documer 3. Copies of the certified copies of the priority documer application from the International Burea * See the attached detailed Office action for a list 	nts have been received. nts have been received in Applicat ority documents have been receiv au (PCT Rule 17.2(a)).	ion No ed in this National Stage
Attachment(s) 1) Notice of References Cited (PTO-892) 2) Notice of Draftsperson's Patent Drawing Review (PTO-948) 3) Information Disclosure Statement(s) (PTO/SB/08) Paper No(s)/Mail Date	4) Interview Summary Paper No(s)/Mail D 5) Notice of Informal F 6) Other:	ate

Application/Control Number: 10/550,137 Page 2

Art Unit: 1641

DETAILED ACTION

Response to Amendment

- 1. Applicant's amendment of claim 25 is acknowledged and has been entered.
- 2. Claims 1, 2, 7-17, 19-21, 24-29 are currently under examination.
- 3. Claim 18 is cancelled.
- 4. Claims 3-6, 22, and 23 are withdrawn.

Rejections Withdrawn

- 5. Applicant's arguments, see p. 7, filed October 27, 2009, with respect to the objection to the drawings have been fully considered and are persuasive. The objection of the drawings has been withdrawn.
- 6. Applicant's arguments, see p. 7 and amended claim 25, filed October 27, 2009, with respect to the rejection of claim 25 under 35 USC 112, 2nd paragraph, have been fully considered and are persuasive. The rejection of claim 25 under 35 USC 112, 2nd paragraph, has been withdrawn.

Claim Rejections - 35 USC § 103

- 7. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:
 - (a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention was made.

Art Unit: 1641

8. Claims 1, 2, 8-17, 19-21, 25-27, and 29 are rejected under 35 U.S.C. 103(a) as being unpatentable over Andersson et al. [US 2003/0053934] in view of Yager et al. [US 2003/0124623].

With respect to claims 1, 2, Andersson et al. teach a microfluidic device comprising a porous matrix placed in a microcavity or immediately downstream a microcavity, a packed bed of monosized particles in the microcavity (para. 0016), wherein the microcavity may further comprise a solid phase with an immobilized affinity reactant (para. 0017). Andersson et al. fails to teach that the porous particles comprise a bed-preserving agent comprising a compound exhibiting a hydrophilic group and is water-soluble.

Yager et al., however, teach a microfluidic device comprising a microfluidic channel and storage area on a wall of the microfluidic channel with a solid reagent plug comprising a matrix, and a reagent comprising particulate materials such as microspheres or nanoparticles having an affinity for binding to the analyte (para. 0011-0017, 0087). Yager et al. further teach that the matrix comprises preservatives such as trehalose and dextran (para. 0089), which are water soluble and exhibit hydrophilic groups. Yager et al. further teach that the once the reagent solutions are rendered solid and dry, the reagents are very robust with respect to storage (para. 0103-0104, 0109).

Therefore, it would have been obvious to one of ordinary skill in the art at the time of the invention to incorporate the matrix of microparticles comprising reagents with the bed-preserving agent of Yager et al. into the device of Andersson et al. in order to be able to render the reactants in the device of Andersson in a solid and dry state, such that the device may be

Application/Control Number: 10/550,137

Art Unit: 1641

stored in conjunction with reagents in a robust manner, thus obviating the need for additional reagents separate from the device.

Page 4

- 9. With respect to claim 8, 9, Yager et al. teach that the microfluidic device may be in a dry state (para. 0104). It is noted that claims 8 and 9 recite product by process limitations. "[E]ven though product-by-process claims are limited by and defined by the process, determination of patentability is based on the product itself. The patentability of a product does not depend on its method of production. If the product in the product-by-process claim is the same as or obvious from a product of the prior art, the claim is unpatentable even though the prior product was made by a different process." *In re Thorpe*, 777 F.2d 695, 698, 227 USPQ 964, 966 (Fed. Cir. 1985).
- 10. With respect to claims 10-11, Andersson et al. teach a microfluidic device comprising a porous matrix placed in a microcavity or immediately downstream a microcavity, a packed bed of monosized particles in the microcavity (para. 0016), wherein the particles would be either swellable or non-swellable.
- 11. With respect to claims 12, 13, Andersson et al. teach an inlet microconduit connected to an inlet port via a volume defining unit that comprises a metering microcavity with valves associated with the outlet ports (para. 0065), wherein the inlet microconduits may simultaneously distribute liquid aliquots to several separate microchannel structures (para. 0066).
- 12. With respect to claim 14, Andersson et al. teaches that a microfluidic device comprising a metering system capable of delivering fluid and passive valves for controlling fluid flow (para. 0065-0068). Therefore, the metering system of Andersson et al. would have sufficient hydrophilicity for being filled by capillarity once an aqueous liquid enters the unit.
- 13. With respect to claim 15, fluid is delivered by centrifugal force (para. 0064).

Art Unit: 1641

- 14. With respect to claims 16, 17, 19-21, Andersson et al. teach that the immobilization of the reactant may comprise streptavidin and a biotinylated affinity reactant (para. 0090), wherein streptavidin would constitute the recited ligand L, and the biotinylated affinity reactant would constitute B-AC_s. Since the binding is between streptavidin and biotin, the affinity constant would not be more than 10³ larger than the affinity constant for streptavidin and biotin.
- 15. With respect to claims 25-27 Andersson et al. teach that the immobilization of the reactant may comprise streptavidin and a biotinylated affinity reactant (para. 0090).
- 16. With respect to claim 29, the antibodies of Yager et al. would have an affinity constant for formation of the complex between the antibody and a desired solute such as an analyte. Since it has been held that where the general conditions of a claim are disclosed in the prior art, discovering the optimum or workable ranged involves only routine skill in the art. *In re Aller*, 105 USPQ 233. Therefore, one of ordinary skill in the art would have found it obvious for the affinity constant for the formation of the complex to be no more than 10⁻⁶ mole/L, through normal optimization procedures known in the art.
- 17. Claims 7, 24, and 28 are rejected under 35 U.S.C. 103(a) as being unpatentable over Andersson et al. [US 2003/0053934] in view of Yager et al. [US 2003/0124623], as applied to claim 1 above, and further in view of Glezer et al. [US 2004/0189311].

With respect to claims 7, 24, 28, Andersson et al. and Yager et al. teach a matrix comprising reagents such as a preservation agent such as trehalose or dextran, but do not teach buffers such as phosphate buffers with a potassium counter-ion.

Art Unit: 1641

Glezer et al., however, teach dry reagent that may include a neutralizing reagent such as phosphate buffering agents (para. 0201), and further teach that this allows for the pH of the extracted sample to be brought to a value that is compatible with subsequent assay reactions carried out on the sample.

Therefore, it would have been obvious to one of ordinary skill in the art at the time of the invention for the reagent of Andersson et al. and Yager et al. to comprise phosphate buffering agents, as suggested by Glezer et al., in order that the extracted samples could be prepared for susbsequent assay reactions and analysis, and the use of a potassium buffer in the device of Andersson et al. would have yielded nothing more than predictable results to one of ordinary skill in the art at the time of the invention.

Double Patenting

18. The nonstatutory double patenting rejection is based on a judicially created doctrine grounded in public policy (a policy reflected in the statute) so as to prevent the unjustified or improper timewise extension of the "right to exclude" granted by a patent and to prevent possible harassment by multiple assignees. A nonstatutory obviousness-type double patenting rejection is appropriate where the conflicting claims are not identical, but at least one examined application claim is not patentably distinct from the reference claim(s) because the examined application claim is either anticipated by, or would have been obvious over, the reference claim(s). See, e.g., *In re Berg*, 140 F.3d 1428, 46 USPQ2d 1226 (Fed. Cir. 1998); *In re Goodman*, 11 F.3d 1046, 29 USPQ2d 2010 (Fed. Cir. 1993); *In re Longi*, 759 F.2d 887, 225 USPQ 645 (Fed. Cir. 1985); *In re Van Ornum*, 686 F.2d 937, 214 USPQ 761 (CCPA 1982); *In re Vogel*, 422 F.2d 438, 164 USPQ 619 (CCPA 1970); and *In re Thorington*, 418 F.2d 528, 163 USPQ 644 (CCPA 1969).

A timely filed terminal disclaimer in compliance with 37 CFR 1.321(c) or 1.321(d) may be used to overcome an actual or provisional rejection based on a nonstatutory double patenting ground provided the conflicting application or patent either is shown to be commonly owned with this application, or claims an invention made as a result of activities undertaken within the scope of a joint research agreement.

Effective January 1, 1994, a registered attorney or agent of record may sign a terminal disclaimer. A terminal disclaimer signed by the assignee must fully comply with 37 CFR 3.73(b).

Application/Control Number: 10/550,137 Page 7

Art Unit: 1641

Claims 1, 8, 16-17, are provisionally rejected on the ground of nonstatutory obviousness-type double patenting as being unpatentable over claims 1-14 of copending Application No. 10/550,182. In particular, the conflicting application recites two or more different sets of microchannel structures comprising a reaction microcavity in which there is a solid phase with an immobilized affinity ligand L directed toward a binder B (claim 1) and wherein the solid phase is in a dry state comprising one or more bed-preserving agents (claim 13). Although the conflicting application does not recite that the reaction microcavity is intended for retaining a solid phase material in the form of a wet porous bed, it would be capable of doing so, and therefore, would render the conflicting application obvious over the instant claims.

This is a <u>provisional</u> obviousness-type double patenting rejection.

Response to Arguments

19. Applicant's arguments filed October 27, 2009 have been fully considered but they are not persuasive. In particular, applicant's arguments that one of ordinary skill in the art is given no guidance by Andersson et al. that a bed-preserving agent is necessary, and therefore there would be no reason to seek resolution of an unrecognized problem. The Office disagrees with this assessment. In particular, if Andersson recognized that a bed-preserving agent would be necessary then the rejection would be an anticipation rejection, as Andersson would have taught all the elements recited in the claim. Because Andersson fails to teach this limitation, Yager is relied upon to teach the missing limitation and to provide motivation for doing so. However, the Office acknowledges that the rejection did not clearly address the motivation for combining the

Application/Control Number: 10/550,137 Page 8

Art Unit: 1641

references, and therefore, the office action has not been made final, in order to allow applicants

to respond to the motivation for combining the references.

Conclusion

20. No claims are allowed.

21. Any inquiry concerning this communication or earlier communications from the

examiner should be directed to Nelson Yang whose telephone number is (571)272-0826. The

examiner can normally be reached on 8:30-5:00.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's

supervisor, Mark Shibuya can be reached on (571)272-0806. The fax phone number for the

organization where this application or proceeding is assigned is 571-273-8300.

22. Information regarding the status of an application may be obtained from the Patent

Application Information Retrieval (PAIR) system. Status information for published applications

may be obtained from either Private PAIR or Public PAIR. Status information for unpublished

applications is available through Private PAIR only. For more information about the PAIR

system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR

system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would

like assistance from a USPTO Customer Service Representative or access to the automated

information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/Nelson Yang/

Primary Examiner, Art Unit 1641